



Sialkot Material Testing Laboratory

UNIT ESTABLISHED BY TRADE DEVELOPMENT AUTHORITY OF PAKISTAN MINISTRY OF COMMERCE GOVT. OF PAKISTAN
NEAR SUBLIME CHOWK, MARALA ROAD, SIALKOT-51310-PAKISTAN
TEL: 0092-52-3562818 simtelstklab@gmail.com FAX: 0092-52-3554217



PD/SIMTEL/2018/392
March 19, 2018

ANALYTICAL/ TEST REPORT

In respect of : M/s. B.A.K. Industry Sialkot
Sample received from : Representative
Name of the Sample : **Hollow Handle Scaler (Point)**
Type/Grade of the Material : AISI-420
Specification against which tested : **Analytical Results Got on Direct Emission Spectrometer**

Results are in percentage:

Cr 12.583	
C 0.187	Mn 0.629
P 0.023	S 0.017
Si 0.269	Ni 0.101

This analytical/test report is based solely on the particular sample(s) supplied by the party and should not be reproduced in part.

Sampling has not been performed by the SIMTEL and the SIMTEL does not accept responsibility that the sample(s) supplied is/are truly representative sample(s) of any batch or stock or entire production.

While the SIMTEL agrees to take every reasonable precaution to ensure validity of its test results, it assumes no liability for the accuracy thereof beyond the amount of the fee charged for the analysis/test.

The party shall assume full responsibility for the ethical use of the results in the analysis/test reports and the laboratory shall be held free from any and all claims which may result from the use of such data by the party or others.

After completion of the report the sample will be preserved for three month until negotiated otherwise.




Engr. Amjad Hussain
Project Director